

Technical Data

R-2A Broth M1687

R-2A Broth is used for cultivation and maintenance of heterotrophic bacteria from potable waters.

Composition**

| Ingredients | Gms / Litre |
|-------------------------|--------------------|
| Casein acid hydrolysate | 0.500 |
| Yeast extract | 0.500 |
| Proteose peptone | 0.500 |
| Dextrose | 0.500 |
| Starch, soluble | 0.500 |
| Dipotassium phosphate | 0.300 |
| Magnesium sulphate | 0.024 |
| Sodium pyruvate | 0.300 |
| Final pH (at 25°C) | 7.2±0.2 |

^{**}Formula adjusted, standardized to suit performance parameters

Directions

Suspend 3.12 grams in 1000 ml distilled water. Heat if necessary to dissolve the medium completely. Dispense into tubes. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. DO NOT OVERHEAT.

Principle And Interpretation

The total bacterial count of drinking water is determined by plate count on a nutritionally rich medium. However all organisms present are not able to grow on them, either because they are slow growers or because they cant grow on that media (1). For this reason a nutritionally reduced medium was described. R-2A Agar is a modification of this medium (2,3).

R-2A Agar is an alternative medium used for the heterotrophic plate counts and for subculturing isolates from potable waters (1). R-2A Agar is also recommended by APHA (4) for pour plate, spread plate and membrane filter technique. R-2A Broth is similar to R-2A Agar except agar. Total count recommended for thebacterial examination of potable waters gives an estimate of the aerobic and facultatively anaerobic bacteria, which grow best at 35°C in a rich medium (3). R-2A Broth enables better recovery of these bacteria from treated waters under different incubation conditions. Many bacteria from natural waters, which contain limited nutrients at ambient temperature, grow best on the media with less nutrient levels. They grow better at the temperatures below the routine laboratory incubation temperatures of 35 to 37°C (3).

This medium contains casein acid hydrolysate, yeast extract, biopeptone as source of essential growth factors required for metabolism of the bacteria. Dextrose is the energy source. Starch acts as a neutralizer that neutralizes any toxic metabolites, if present. Phosphate buffers the medium while sodium pyruvate supplies additional nutrition. Magnesium sulphate serves as a source of ions. Due to the presence of the above mentioned ingredients these media allow the growth of stressed and chlorine tolerant bacteria present in treated waters.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Colour and Clarity of prepared medium

Yellow coloured, clear solution in tubes

Reaction

Reaction of 0.312% w/v aqueous solution at 25°C. pH: 7.2±0.2

pН

7.00-7.40

Cultural Response

HiMedia Laboratories Technical Data

M1687: Cultural characteristics observed *by using standard ATCC cultures after an incubation at 35-37°C for 24-72 hours. (*-In case of water samples from fields it is suggested to incubate further for upto 7 days to ascertain the absence of organisms)

| Organism | Inoculum (CFU) | Growth |
|--------------------------------------|-------------------|----------------|
| Cultural Response | | |
| Candida albicans ATCC 10231 | 50-100 | good-luxuriant |
| Enterococcus faecalis ATCC 29212 | 50-100 | good-luxuriant |
| Escherichia coli ATCC 25922 | 50-100 | good-luxuriant |
| Salmonella Enteritidis ATCC 13076 | '50-100 | good-luxuriant |
| Salmonella Typhi ATCC 6539 | 50-100 | good-luxuriant |

Storage and Shelf Life

Store below 30°C in tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label.

Reference

- 1.Reasoner and Geldreich, 1985, Appl. Environ. Microbiol., 49:1. 2.Stark and McCoy. 1938. Zentralbl. Bacteriol. Parasitenkd. Infectionskr. Hyg. Abt.2 98: 201
- 3. Collins and Willoughby, 1962, Arch. Microbiol., 43:294.
- 4.Greenberg A. E., Trussell R. R. and Clesceri L. S. (Eds.), 1985, Standard Methods for the Examination of Water and Wastewater, 16th ed., APHA, Washington, DC

Revision: 2 / 2015

Disclaimer:

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMediaTM publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMediaTM Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.