



Diagnostic Sensitivity Test Agar (D.S.T. Agar)

M502

Intended Use:

Diagnostic Sensitivity Test Agar (D.S.T. Agar) is used as an antibiotic sensitivity-testing medium for antibiotic sensitivity testing of fastidious pathogens such as *Neisseria*, *Streptococcus* and *Haemophilus* species with blood enrichment from clinical samples.

Composition**

Ingredients	Gms / Litre
Proteose peptone	10.000
HMV infusion solids #	10.000
Dextrose (Glucose)	2.000
Sodium chloride	3.000
Disodium phosphate	2.000
Sodium acetate	1.000
Adenine sulphate	0.010
Guanine hydrochloride	0.010
Uracil	0.010
Xanthine	0.010
Aneurine	0.00002
Agar	15.000
Final pH (at 25°C)	7.4±0.2

**Formula adjusted, standardized to suit performance parameters

Equivalent to Veal infusion solids

Directions

Suspend 43.04 grams in 1000 ml distilled water. Heat to boiling to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. For blood agar, cool the base to 45-50°C and add 7% v/v sterile defibrinated horse blood aseptically. Mix well with gentle rotation and pour into sterile Petri plates.

Principle And Interpretation

Diagnostic Sensitivity Test Agar is recommended for diagnostic as well as testing susceptibility of organisms to antibiotics and chemotherapeutic agents such as Sulphonamides. The latter produce well defined zones due to the absence of interfering substances.

The medium is nutritionally rich due to presence of amino acid bases and glucose. The salts present, helps in avoiding sudden pH shifts due to acid production, which might affect the susceptibility test and haemolytic reactions (1) and the MIC values of pH susceptible antimicrobials (2). Aneurine acts as vitamin source which improves the growth of several organisms especially Staphylococci. The agar used in the formulation has been specially processed to allow unimpeded diffusion of antimicrobials from discs (3). Addition of the bases like adenine, guanine, uracil and xanthine improve the antibiotic testing performance of the medium.

The reactive levels of thymidine and thymine must be sufficiently reduced to avoid antagonism of trimethoprim and sulphonamides which is an essential requirement for satisfactory antimicrobial susceptibility media. The requirement is achieved by addition of lysed horse blood to Diagnostic Sensitivity Testing medium. The level of thymidine is further reduced due to the action of thymidine phosphorylase, released from lysed horse erythrocytes (4). Thymidine-dependant organisms will not grow in absence of thymidine or will grow poorly in media containing reduced levels (5).

For less demanding organisms like Micrococci, *Salmonella*, *Shigella* coliform bacteria and *Proteus* species, this medium can be used without blood. For fastidious organisms like *Haemophilus influenzae*, *Neisseria meningitides*, alpha and beta haemolytic Streptococci blood enrichment is necessary.

Antibiotic susceptibility test is performed as follows: Suspension of test organisms is spread on the surface of the medium. Sensitivity discs (3) are equally spaced on the seeded medium surface and incubated at 37°C for 18 hours. The zones of inhibition obtained are recorded. This medium has reduced thymidine activity and this will affect its performance as a

primary isolation medium.

Type of specimen

Clinical samples : Pure samples from Blood, urine, respiratory samples, and other clinical material

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (8,9).

Warning and Precautions

In Vitro diagnostic use only. Read the label before opening the container. Wear protective gloves/protective clothing/ eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

- 1.The salts present, helps in avoiding sudden pH shifts due to acid production, which might affect the susceptibility test and haemolytic reactions (1) and the MIC values of pH susceptible antimicrobials (2).
2. Some fastidious organisms may not grow.
3. Inoculum density affects inhibition zone. Heavy inoculum may result in smaller zones while scanty growth may result in enlarged zones.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Gelling

Firm, comparable with 1.5% Agar gel

Colour and Clarity of prepared medium

Basal medium : Light amber coloured, clear to slightly opalescent gel forms. After addition of 7%w/v sterile defibrinated blood : Cherry red coloured, opaque gel forms in Petri plates.

Reaction

Reaction of 4.3% w/v aqueous solution at 25°C. pH : 7.4±0.2

pH

7.20-7.60

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours.

Cultural Response

Organism	Growth	Inoculum (CFU)	Recovery
Cultural Response <i>Staphylococcus aureus</i> ATCC 25923 (00034*)	luxuriant	50-100	≥70%
<i>Escherichia coli</i> ATCC 25922 (00013*)	luxuriant	50-100	≥70%
<i>Enterococcus faecalis</i> ATCC 29212 (00087*)	luxuriant	50-100	≥70%
<i>Neisseria meningitidis</i> ATCC 13090	luxuriant	50-100	≥70%

<i>Proteus mirabilis</i> ATCC 25933	luxuriant	50-100	>=70%
<i>Micrococcus luteus</i> ATCC 10240	luxuriant (with the addition of blood)	50-100	>=70%
<i>Salmonella Typhi</i> ATCC 6539	luxuriant	50-100	>=70%
<i>Streptococcus pneumoniae</i> ATCC 6305	luxuriant(with the addition of blood)	50-100	>=70%
<i>Streptococcus pyogenes</i> ATCC 19615	Luxuriant (with the addition of blood)	50-100	>=70%
<i>Shigella flexneri</i> ATCC 12022 (00126*)	luxuriant	50-100	>=70%

Key : *Corresponding WDCM numbers.

Storage and Shelf Life

Store below 10-30°C in a tightly closed container and the prepared medium at 2-8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (8,9).

Reference

1. Expert Committee on antibiotics, 1961, World Health Organisation Technical Report Series No. 210, WHO, Geneva.
2. Bechtle R. M. and Schere G. H., 1958, Antibiotics and Chemotherapy, 8(12): 599.
3. Marshall J. H. and Kelsey J. C., 1960, J. Hyg. Camb., 58 : 367.
4. Ferone R., Bushby S. R. M., Burchall J. J., Moore W. D., and Smith D., 1975, Antimicrobial Agents Chemotherap., 7: 91-98.
5. Downes F. P. and Ito K. (Ed.), 2001, Compendium of Methods for the Microbiological Examination of Foods, 4th ed., APHA, Washington, D.C.
6. Rice E.W., Baird, R.B., Eaton A. D., Clesceri L. S. (Eds.), 2012, Standard Methods for the Examination of Water and Wastewater, 22nd ed., APHA, Washington, D.C.
7. Wehr H. M. and Frank J. H., 2004, Standard Methods for the Microbiological Examination of Dairy Products, 17th Ed., APHA Inc., Washington, D.C.
8. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
9. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.


Revision : 03/ 2018



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg, Mumbai-86, MS, India



CE Partner 4U, Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.