

<i>Salmonella Typhimurium</i> ATCC 14028 (00031*)	50-100	good-luxuriant	>=50%	black with metallic sheen
<i>Shigella flexneri</i> ATCC 12022 (00126*)	50-100	none-poor	<=10%	brown
<i>Escherichia coli</i> ATCC 8739 (00012*)	50-100	none-poor	<=10%	brown to green, depends on inoculum density
<i>Escherichia coli</i> NCTC 9002	50-100	none-poor	<=10%	brown to green (depends on inoculum density)
<i>Salmonella Abony</i> NCTC 6017 (00029*)	50-100	good-luxuriant	>=50%	black with metallic sheen

Key : *Corresponding WDCM numbers.

#- Formerly known as *Enterobacter aerogenes*

Storage and Shelf Life

Store below 30°C in a tightly closed container and the prepared medium at 2 - 8°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (16,17).

Reference


- Tindall B. J., Crimont P. A. D., Gorrity G. M., EUZESY B. P., 2005, Int. J. Sys. Evol. Microbiol., 55:521
- Koneman E. W., Allen S. D., Janda W. M., Schreckenberger P. C., Winn W. C. Jr., 1992, Colour Atlas and Textbook of Diagnostic Microbiology, 4 th Ed., J. B. Lippincott Company
- Mandell G. L., Douglas R. G. Jr., Bennet J. E., (Eds.), 1985, Principles and Practice of Infectious Diseases, 2nd Ed., 660-669, John Wiley & Sons New York.
- Gunter and Tuft, 1939, J. Lab. Clin. Med., 24:461.
- Wilson and Blair, 1926, J. Pathol. Bacteriol., 29:310.
- Wilson and Blair, 1927, J. Hyg., 26:374
- Wilson and Blair, 1931, J. Hyg., 31:138
- Washington J. A., 1981, Laboratory Procedures in Clinical Microbiology, Springer-Verlag, New York.
- gpl, *C rml ,B,* lb Pgac C,U,* (Eds.), 20/5, Standard Methods for the Examination of Water and Wastewater, 23rd ed., APHA, Washington, D.C.
- FDA Bacteriological Analytical Manual, 2005, 18th Ed., AOAC, Washington, D.C.
- Murray P. R., Baron J. H., Pfaller M. A., Tenover F. C. and Tenover R. H., (Eds.). 1999, Manual of Clinical Microbiology, 7th Ed., American Society for Microbiology, Washington, D.C.
- Salfinger Y., and Tortorello M.L. Fifth (Ed.), 2015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.
- Indian Pharmacopoeia, 1996, Ministry of Health and Family Welfare, Govt. of India, Volume 2.
- MacFaddin J. F., 2000, (Ed.), Biochemical Tests for Identification of Medical Bacteria, 3rd Edition, Lippincott, Williams & Wilkins, New York.
- American Public Health Association, Standard Methods for the Examination of Dairy Products, 1978, 14th Ed., Washington D.C.
- Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
- Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.



In vitro diagnostic medical device



CE Marking




Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg, Mumbai-86, MS, India



CE Partner 4U ,Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.