



M-E.coli Broth

M1426

Intended Use:

Recommended for the detection, differentiation and enumeration of *Escherichia coli* and coliforms in water samples using membrane filter technique without further confirmation. It can also be used to isolate *E.coli* from clinical samples.

Composition**

Ingredients	Gms / Litre
Tryptone	20.000
Bile salts mixture	1.500
Chromogenic mixture	0.175
Final pH (at 25°C)	7.2±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 21.67 grams in 1000 ml distilled water. Heat if necessary to dissolve the medium completely. DO NOT AUTOCLAVE. Cool to 45-50°C, aseptically add desired quantity (2 to 5 ml) of broth on sterile absorbent cotton pad for saturation. The medium should be used within 24 hours of rehydration.

Principle And Interpretation

M-E.coli Broth is used for detection and differentiation of *Escherichia coli* and coliforms in water samples using membrane filter technique. It is based on Tryptone Bile Agar used for detection of *Escherichia coli* in foods (1) where recovery of *Escherichia coli* is faster, more reliable and accurate.

The water sample is filtered through membranes and then placed on pad saturated with M-E.coli Broth and incubated at 37°C in sealed Petri plates. The medium contains chromogenic mixture, which helps to detect glucuronidase activity of *Escherichia coli* (3). This specific enzyme differentiates *Escherichia coli* from other coliforms. *Escherichia coli* cells split the chromogenic mixture with the help of glucuronidase to give blue colouration to the colonies. Coliforms other than *Escherichia coli* turn red as they reduce TTC (2,3,5-triphenyl tetrazolium chloride). Thus, the resulting colour distinction allows simple interpretation of test without further confirmation. tryptone provides the essential growth nutrients to the organisms. Bile salts inhibit gram - positive organisms.

Type of specimen

Clinical samples- faeces, urine ; Water samples

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (4,5).

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards.(2)

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions

In Vitro diagnostic use only. Read the label before opening the container. Wear protective gloves/protective clothing/ eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations

1. β -glucuronidase is present in 97% of *E.coli* strains, however few *E.coli* may be negative. 2. Some species may show poor growth due to nutritional variations.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Quality Control

Appearance

Light yellow to beige homogeneous free flowing powder

Colour and Clarity of prepared medium

Light yellow coloured, clear solution without any precipitate

Reaction

Reaction of 2.16 % w/v aqueous solution at 25°C. pH : 7.2±0.2

pH

7.00-7.40

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours

Organism	Inoculum (CFU)	Growth	Colour of Colony
<i>Enterobacter aerogenes</i> ATCC 13048 (00175*)	50-100	luxuriant	red
<i>Escherichia coli</i> ATCC 25922 (00013*)	50-100	luxuriant	blue
<i>Staphylococcus aureus</i> ATCC 25923 (00034*)	$\geq 10^3$	inhibited	

Key : *Corresponding WDCM numbers.

Storage and Shelf Life

Store dehydrated powder and prepared medium on receipt at 2-8°C. Use before expiry period on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (4,5).

Reference

1. Anderson J. M. and Baird Parker A.C., (1975), J. Appl. Bact., 39:111.
2. Hansen W. and Yourassawsky E., (1984), J. Clin. Microbiol. 20:1177.
3. Baird R.B., Eaton A.D., and Rice E.W., (Eds.), 2015, Standard Methods for the Examination of Water and Wastewater, 23rd ed., APHA, Washington, D.C.
4. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
5. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.


Revision : 02 / 2018



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg, Mumbai-86, MS, India



CE Partner 4U ,Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.