



Alkaline Peptone Water

M618

Intended use

Alkaline Peptone Water is recommended for enrichment of *Vibrio* species from food, water and clinical samples.

Composition**

Ingredients	Gms / Litre
Peptone	10.000
Sodium chloride	10.000
Final pH (at 25°C)	8.4±0.2

**Formula adjusted, standardized to suit performance parameters

Directions

Suspend 20.0 grams in 1000 ml purified/distilled water. Heat if necessary to dissolve the medium completely. Dispense in tubes or flasks as desired. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes.

Principle And Interpretation

Clinical materials containing small numbers of *Vibrio* should be inoculated into an enrichment medium prior to plating onto a selective medium, such as TCBS Agar (M189). Alkaline Peptone Water is a suitable enrichment broth for this purpose (4-6).

The relatively high pH of the medium (approximately 8.4) provides a favorable environment for the growth of *Vibrio*'s.

This medium is recommended by APHA (8) for enrichment of *Vibrio* species from seafood, infectious materials and other clinical specimens such as faeces (2).

Peptone provides nitrogen and carbon source, long chain amino acids, vitamins and other essential nutrients. Sodium chloride maintains osmotic equilibrium.

Add 10 grams of seafood to 90 ml of Alkaline Peptone Water and incubate for upto 18-20 hours at 37°C. Prolonged incubation will result in growth of the suppressed contaminating organisms to develop (3). Growth in tubes is indicated by turbidity compared to an un-inoculated tube (control). Growth from the enrichment broth is used for plating on selective media. For biochemical identification a pure culture is recommended.

Type of specimen

Clinical samples: faeces; Food samples ; Water samples

Specimen Collection and Handling

For clinical samples follow appropriate techniques for handling specimens as per established guidelines (6,7).

For food samples, follow appropriate techniques for sample collection and processing as per guidelines (8).

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standards (1).

After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions :

In Vitro diagnostic Use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/ face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets

Limitations :

1. Certain strains of *Vibrio* species requiring higher sodium chloride concentration may show poor growth.
2. Further recovery from this enriched broth onto selective media is required.
3. Biochemical characterization is carried out from pure isolates for complete identification.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the recommended temperature.

Quality Control

Appearance

Cream to yellow homogeneous free flowing powder

Colour and Clarity of prepared medium

Light yellow coloured clear solution without any precipitate

Reaction

Reaction of 2% w/v aqueous solution at 25°C. pH : 8.4±0.2

pH

8.20-8.60

Cultural Response

M618: Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours.

Organism	Inoculum (CFU)	Growth
Cultural Response		
<i>Vibrio cholerae</i> ATCC 15748	50-100	luxuriant
<i>Vibrio parahaemolyticus</i> ATCC 17802 (00037*)	50-100	luxuriant

Key : (*) Corresponding WDCM numbers.

Storage and Shelf Life

Store between 10-30°C in a tightly closed container and the prepared medium at 15-25°C. Use before expiry date on the label. On opening, product should be properly stored dry, after tightly capping the bottle in order to prevent lump formation due to the hygroscopic nature of the product. Improper storage of the product may lead to lump formation. Store in dry ventilated area protected from extremes of temperature and sources of ignition Seal the container tightly after use. Use before expiry date on the label.

Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (6,7).

Reference

- Baird R.B., Eaton A.D., and Rice E.W., (Eds.), 2015, Standard Methods for the Examination of Water and Wastewater, 23rd ed., APHA, Washington, D.C.
- Cruikshank R., 1968, Medical Microbiol., 11th Ed., Livingstone Ltd., London
- Finegold S. M. and Martin W. J., 1982, W. J. Bailey and Scotts Diagnostic Microbiol, 6th Ed., C.V. Mosby Co., St. Louis, p. 242
- Forbes B. A., Sahm A. S., and Weissfeld D. F., Bailey & Scotts Diagnostic Microbiology, 10th Ed., 1998, Mosby, Inc., St. Louis, Mo.
- Gilligan, Janda, Karmali and Miller, 1992, Cumitech 12A, Laboratory Diagnosis of Bacterial Diarrhea, Coord. Ed., Nolte, American Society for Microbiology, Washington, D.C.
- Isenberg, (Ed.), 1992, Clinical Microbiology Procedures Handbook, Vol. I, American Society for Microbiology, Washington, D.C.
- Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock, D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
- Salinger Y., and Tortorello M.L. Fifth (Ed.), 2015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.


Revision: 03 / 2018



In vitro diagnostic medical device



CE Marking



Storage temperature



Do not use if package is damaged



HiMedia Laboratories Pvt. Limited,
23 Vadhani Industrial Estate,
LBS Marg, Mumbai-86, MS, India



CE Partner 4U ,Esdoornlaan 13, 3951
DB Maarn The Netherlands,
www.cepartner4u.eu

Disclaimer :

User must ensure suitability of the product(s) in their application prior to use. Products conform solely to the information contained in this and other related HiMedia™ publications. The information contained in this publication is based on our research and development work and is to the best of our knowledge true and accurate. HiMedia™ Laboratories Pvt Ltd reserves the right to make changes to specifications and information related to the products at any time. Products are not intended for human or animal or therapeutic use but for laboratory, diagnostic, research or further manufacturing use only, unless otherwise specified. Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.